

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/22/2014

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:43
INTERNAL STANDARD	VERIFIED	15:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:44
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	15:46
BLANK TEST	0.00	15:47

*J.M. Lopez #299*

