

SUPERVISOR TEST




WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208

Date: 09/23/2014

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00



*See By Report WSC #142*  
 09-23-14