SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/15/2014

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:41
INTERNAL STANDARD	VERIFIED	17:41
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	17:42
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	17:43
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	17:44
BLANK TEST	0.00	17:45

P.O. # 301
10/15/14