

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/27/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00



SOS *Ralph Bartlett* WESO #102
10-27-14