

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/10/2014

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:49
INTERNAL STANDARD	VERIFIED	16:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:49
BLANK TEST	0.00	16:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:51
BLANK TEST	0.00	16:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:52
BLANK TEST	0.00	16:53

