

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/30/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36

Sgt. Ralph Bartlett
weso #162 12-30-14

