

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 03/31/2014

OPERATOR NAME:  
JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:21
INTERNAL STANDARD	VERIFIED	15:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:22
BLANK TEST	0.00	15:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:23
BLANK TEST	0.00	15:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25

*JM Lopez 299*

