

SOLUTION CHANGE



WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/28/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:05
INTERNAL STANDARD	VERIFIED	03:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:06
BLANK TEST	0.00	03:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10



SGS By [Signature] 6010 #162

04-28-14