

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106606  
 -----

Date: 05/19/2014

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:33
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:36
BLANK TEST	0.00	22:37



SGT Rly Bartlett WPSO #162

05-19-14