

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/01/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:46
INTERNAL STANDARD	VERIFIED	03:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:47
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:48
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:49
BLANK TEST	0.00	03:50



SG- Rly Printer w/c #162

07-01-14