

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/07/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:49
BLANK TEST	0.00	02:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:53



SGT R. J. Bartlett

WC50 #162

07-07-14