

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/15/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:10
BLANK TEST	0.00	03:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:12
BLANK TEST	0.00	03:13



SGT Ralph J Bartlett WRSO #162

07-15-14