

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 07/29/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:08
INTERNAL STANDARD	VERIFIED	03:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12

SGS *RJ Bartlett* 07-29-14

