

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/04/2014

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:38
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:40
BLANK TEST	0.00	02:41



SGT R.J. Bartlett WSO # 102 08-04-14