

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/22/2014

OPERATOR NAME:
JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:22
INTERNAL STANDARD	VERIFIED	18:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:23
BLANK TEST	0.00	18:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:24
BLANK TEST	0.00	18:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:26
BLANK TEST	0.00	18:26

JM Lopez #297

