

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 09/09/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:24
BLANK TEST	0.00	03:25



Sgt Ralph Bartlett WESO #162 09-09-14