

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 11/03/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:19
INTERNAL STANDARD	VERIFIED	17:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:19
BLANK TEST	0.00	17:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:20
BLANK TEST	0.00	17:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:22
BLANK TEST	0.00	17:22



SGT Ray [Signature] WCSO #162
11-3-14