

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/02/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:54
INTERNAL STANDARD	VERIFIED	03:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:54
BLANK TEST	0.00	03:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:56
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:57
BLANK TEST	0.00	03:58



SGS R/L / *[Signature]* WESU #102
12-02-14