

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/08/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:17
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:18
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:20
BLANK TEST	0.00	02:21



set by *R. Bartlett* WOSO #12

12-08-14