

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 12/22/2014

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:57
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:00
BLANK TEST	0.00	01:01



SOT *Ralph Bartlett* Wesc #16
 12-22-14