

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/30/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:55
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:56
BLANK TEST	0.00	02:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	02:57
BLANK TEST	0.00	02:58



Sgt Ralph Bartlett
WCSO # 162 12-30-14