

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/14/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:22
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25

