

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 01/28/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:14
INTERNAL STANDARD	VERIFIED	01:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17



565 *RJ Bartlett*

WCSO #162

01-28-14