

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/03/2014

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:53
INTERNAL STANDARD	VERIFIED	00:53
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56



SGS RAJ B... WSC #102 02-03-14