

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/17/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:02
BLANK TEST	0.00	01:03



SGS Rep of Bartlett Warren Sheriff #162 02-17-14