

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/25/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:04
INTERNAL STANDARD	VERIFIED	02:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08



SG- RA, Bartlett wesc #162 02-25-14