

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/11/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:23
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27



SG *Ralph Bartlett* was off 162

03-11-14