

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 03/25/2014

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 13320

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:11
INTERNAL STANDARD	VERIFIED	01:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15



SGS By J. Bartlett

WC30 #162

03-25-14