

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/01/2014

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:02
BLANK TEST	0.00	10:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:03
BLANK TEST	0.00	10:03

J. Lopez 299

