

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/07/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:06
INTERNAL STANDARD	VERIFIED	13:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:07
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	13:09
BLANK TEST	0.00	13:10



565 R/J Bartlett WCSO #162 0407-14