

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/14/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:41
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:44



SOS RJA Bartlett

WOB #102

04-14-14