

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/22/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:27
INTERNAL STANDARD	VERIFIED	22:27
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:28
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:29
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:30
BLANK TEST	0.00	22:31



SG- RJB

WCSO #162

04-22-14