

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/28/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:00
INTERNAL STANDARD	VERIFIED	02:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04



565 *RJ Bartlett* WESO #162 04-28-14