

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 05/15/2014

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:07
INTERNAL STANDARD	VERIFIED	04:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:07
BLANK TEST	0.00	04:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:08
BLANK TEST	0.00	04:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:09
BLANK TEST	0.00	04:10



SG- Rly Rmney WCSO #142 05-15-14