

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/03/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:59
INTERNAL STANDARD	VERIFIED	01:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02



SGS *Ralph Bartlett* W90 #162 06-07-14