

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/09/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:50
INTERNAL STANDARD	VERIFIED	03:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:52
BLANK TEST	0.00	03:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54



SGT R. J. Bartlett WCSO #162
06-09-14