

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/17/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:23
INTERNAL STANDARD	VERIFIED	01:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27



SGT RJA / Bartlett WCSO #162 06-17-14