

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/23/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:12
INTERNAL STANDARD	VERIFIED	01:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16



SGT *Ralph Bartlett* WCO #161 06-23-14