

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/01/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14080

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 02:48 |
| INTERNAL STANDARD       | VERIFIED | 02:49 |
| REFERENCE STANDARD TEMP | 33.97c   |       |
| REFERENCE STANDARD      | 0.09     | 02:49 |
| BLANK TEST              | 0.00     | 02:50 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 02:50 |
| BLANK TEST              | 0.00     | 02:51 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 02:51 |
| BLANK TEST              | 0.00     | 02:52 |



SGS RJB Bartlett WCSO# 122 07-01-14