

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/07/2014

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14080

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:30
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34



565 Rly Dwyer WESU #162

07-07-14