

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 07/29/2014

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:52
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:55



505 *Ralph Bartlett* 07-29-14