

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/12/2014

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53



SCS *Ralph Bartlett* W#50 #162 08-11-14