

SOLUTION CHANGE

-----  
 WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 08/22/2014

OPERATOR NAME:  
 JOSHUA M LOPEZ  
 Permit #: 34728  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14140

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 16:42 |
| INTERNAL STANDARD       | VERIFIED | 16:42 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 16:42 |
| BLANK TEST              | 0.00     | 16:43 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 16:43 |
| BLANK TEST              | 0.00     | 16:44 |
| REFERENCE STANDARD TEMP | 33.97c   |       |
| REFERENCE STANDARD      | 0.09     | 16:44 |
| BLANK TEST              | 0.00     | 16:45 |

*Joshua M Lopez 2/19*

