

SOLUTION CHANGE




WARREN COUNTY
 SHERRIF'S OFFICE
 DATAMASTER dmt: 144506

Date: 09/01/2014

OPERATOR NAME:
 RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:33
INTERNAL STANDARD	VERIFIED	01:33
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:37



SGS *RJ Bartlett* WES# 162
 09-01-14