

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/09/2014

OPERATOR NAME:
RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:48



SOS Ralph Bartlett W.C. SO #162 09-09-14