

SOLUTION CHANGE


WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/29/2014

OPERATOR NAME:  
RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:40
INTERNAL STANDARD	VERIFIED	04:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:44



SGT Barty Bartlett W/SO # 162  
09-29-14