

SUPERVISOR TEST




WARREN COUNTY
SHERRIF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/15/2014

OPERATOR NAME:
KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:14
INTERNAL STANDARD	VERIFIED	18:14
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	18:15
BLANK TEST	0.00	18:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:16
BLANK TEST	0.00	18:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:17
BLANK TEST	0.00	18:18



RO. [Signature] 301
10/15/14