

SOLUTION CHANGE




WARREN COUNTY  
 SHERRIF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 11/03/2014

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14180

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 17:50 |
| INTERNAL STANDARD       | VERIFIED | 17:50 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 17:50 |
| BLANK TEST              | 0.00     | 17:51 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 17:51 |
| BLANK TEST              | 0.00     | 17:52 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 17:53 |
| BLANK TEST              | 0.00     | 17:53 |



Sus Ralph Bartlett USSO #162  
 11-3-14