

SUPERVISOR TEST

-----  
WARREN COUNTY  
SHERRIF'S OFFICE  
DATAMASTER dmt: 144506  
-----

Date: 11/16/2014

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 14180

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:22
INTERNAL STANDARD	VERIFIED	09:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:23
BLANK TEST	0.00	09:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:24
BLANK TEST	0.00	09:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:25
BLANK TEST	0.00	09:26

