

SOLUTION CHANGE

WARREN COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506

Date: 12/02/2014

OPERATOR NAME:  
 RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 14230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:54
INTERNAL STANDARD	VERIFIED	01:54
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:55
BLANK TEST	0.00	01:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:58

SG: *RJ Bartlett* WCO #142  
 12-02-14

